



Antibody and Immunoassay Services

Li Ka Shing Faculty of Medicine, The University of Hong Kong

Rapid Human Cystatin C Immunoassay Kit

Catalogue Number: 31241

For the quantitative determination of human cystatin C
concentrations in serum, plasma and urine samples

This package insert must be read in its entirety before using this product
Use only the current version of product data sheet enclosed with the kit

L8-37, Laboratory Block, The University of Hong Kong,
21 Sassoon Road, Hong Kong
Tel: ++852-2819-9743 Fax: ++852-2816-2095
Website: www.antibody.hku.hk e-mail: antibody@hku.hk

TABLE OF CONTENTS

Contents	Page
INTRODUCTION	1
PRINCIPLE OF THE ASSAY	1
INTENDED USE	2
REAGENTS SUPPLIED	2
OTHER MATERIALS REQUIRED, BUT NOT PROVIDED	2
STORAGE	3
PREPARATION OF REAGENTS	3
PREPARATION OF STANDARDS AND SAMPLES	4
ASSAY PROCEDURE	5
CALCULATION	5
TYPICAL STANDARD CURVE	6
ASSAY CHARACTERISTICS	7
REFERENCES	7
SUMMARY OF ASSAY PROCEDURE	8

INTRODUCTION

Human cystatin c (or cystatin 3), which is composed of 120 amino acid residues, belongs to the cystatins superfamily that inactivates lysosomal cysteine proteinases. As a strongly cationic and low-molecular weight (13.4 kDa) protein, it is almost freely filtered across the glomerular membrane, and is mainly used as a biomarker of kidney function. A growing body of evidence suggests that cystatin c is a more reliable biomarker of glomerular filtration rate than creatinine ^[1-3].

In addition to kidney disease, altered serum levels of cystatin c are associated with several types of cardiovascular disease, including myocardial infarction, stroke, heart failure, peripheral arterial disease and metabolic syndrome ^[4-7]. It also seems to play a role in brain disorders involving amyloid, such as Alzheimer's disease ^[8, 9]. Furthermore, Cystatin c has also been investigated as a prognostic marker in several forms of cancer ^[11, 12].

PRINCIPLE OF THE ASSAY

This assay is a sandwich ELISA designed for the quantitative detection of human cystatin c in samples in 1 hour. The immunoplate is pre-coated with antibody specific to human cystatin c. Standards and samples are pipetted into the wells and any human cystatin c present is sandwiched by the immobilised antibody and a second horseradish peroxidase (HRP)-linked antibody specific to human cystatin c that is co-incubated with the samples. After wash step to remove any unbound substances, the HRP substrate solution is added and colour develops in proportion to the amount of human cystatin c bound initially. The assay is stopped and the optical density of the wells determined using a micro-plate reader. Since the increases in absorbance are directly proportional to the amount of captured human cystatin c, the unknown sample concentration can be interpolated from a reference curve included in each assay.

INTENDED USE

This Rapid Human Cystatin C ELISA kit is designed for quantification of human cystatin c in serum, plasma and urine samples.

REAGENTS SUPPLIED

Each kit is sufficient for one 96-well plate and contains the following components:

1. Micro-titre strips (96 wells)-Coated with antibody against human cystatin c.
2. 10×Wash buffer-30 ml.
3. 5×Assay buffer-30 ml.
4. 100×Detection antibody solution-Antibody against human cystatin c conjugated with horseradish peroxidase, 0.12 ml.
5. Human cystatin c standard-15 ng of native human cystatin c in a buffered protein base, lyophilised.
6. Substrate solution-12 ml, ready for use.
7. Stop solution-12 ml, ready for use.
8. Plate cover- 1.

OTHER MATERIALS REQUIRED, BUT NOT PROVIDED

1. Pipettes and pipette tips.
2. 96-well plate or manual strip washer.
3. Buffer and reagent reservoirs.
4. Paper towels or absorbent paper.
5. Plate reader capable of reading absorbency at 450 nm.
6. Distilled water or deionised water.
7. Horizontal micro-plate shaker capable of 600 rpm.

STORAGE

The kit should be stored at 2-8°C upon receipt, and all reagents should be equilibrated to room temperature before use. Remove any unused antibody-coated strips from the Human cystatin c micro-plate, return them to the foil pouch and re-seal. Once opened, the strips may be stored at 2-8°C for up to one month.

PREPARATION OF REAGENTS

Bring all reagents and materials to room temperature before assay.

A. 1×Assay buffer.

Prepare 1×Assay buffer by mixing the 5×Assay buffer (30 ml) with 120 ml of distilled water or deionised water. If precipitates are observed in the 5×Assay buffer bottle, warm the bottle in a 37°C water bath until the precipitates disappear. The 1×Assay buffer may be stored at 2-8°C for up to one month.

B. 1×Wash buffer.

Prepare 1×Wash buffer by mixing the 10×Wash buffer (30 ml) with 270 ml of distilled water or deionised water. If precipitates are observed in the 10×Wash buffer bottle, warm the bottle in a 37°C water bath until the precipitates disappear. The 1×Wash buffer may be stored at 2-8°C for up to one month.

C. 1×Detection antibody solution.

Spin down the 100×Detection antibody solution briefly and dilute the desired amount of the antibody 1:100 with 1×Assay buffer, 100 µl of the 1×Detection antibody solution is required per well. Prepare only as much 1×Detection antibody solution as needed. Return the 100×Detection antibody solution to 2-8°C immediately after the necessary volume is removed.

PREPARATION OF STANDARDS AND SAMPLES

Human cystatin c standards: Reconstitute the lyophilised standard with 1 ml of 1×Assay buffer to generate a standard stock solution of 15 ng/ml. Allow the standard to sit for 10 minutes with gentle agitation prior to making dilutions. Prepare serially diluted standards using 1×Assay buffer as follows:

Standard volume	Volume of 1×Assay buffer	Concentration
15 ng/ml stock	-	15 ng/ml
250 µl of 15 ng/ml stock	250 µl	7.5 ng/ml
250 µl of 7.5 ng/ml std.	250 µl	3.75 ng/ml
250 µl of 3.75 ng/ml std.	250 µl	1.87 ng/ml
250 µl of 1.87 ng/ml std.	250 µl	0.93 ng/ml
250 µl of 0.93 ng/ml std.	250 µl	0.46 ng/ml
250 µl of 0.46 ng/ml std.	250 µl	0.23 ng/ml

1×Assay buffer serves as the zero standard (0 ng/ml). The reconstituted standard stock should be aliquoted and frozen at -20°C for one month. Avoid repeating freezing/thawing cycles. Please do not store the diluted standard solutions.

Sample preparation

Serum or plasma sample is generally required a 160-fold dilution in this assay. A suggested dilution step is to add 5 µl of serum or plasma sample to 795 µl of 1×Assay buffer. Urine sample is generally required a 20-fold dilution. A suggested dilution step is to add 50 µl of urine sample to 950 µl of 1×Assay buffer.

ASSAY PROCEDURE

It is recommended that all standards and samples be assayed in duplicate.

1. Add 100 μ l of standard or sample per well.
2. Add 100 μ l of the 1 \times Detection antibody solution to each well, seal the plate with a plate cover. Incubate at room temperature for 30 minutes, shaking the plate at 600 rpm on a horizontal micro-plate shaker.
3. Discard the content and tap the plate on a clean paper towel to remove residual solution in each well. Add 300 μ l of 1 \times Wash buffer to each well and incubate for 1 minute. Discard the 1 \times Wash buffer and tap the plate on a clean paper towel to remove residual wash buffer. Repeat the wash step for a total 4 washes.
4. Add 100 μ l of Substrate solution to each well. Incubate at room temperature for 15 minutes. **Protect from light.**
5. Add 100 μ l of Stop solution to each well, gently tap the plate frame for a few seconds to ensure thorough mixing.
6. Measure absorbance of each well at 450 nm immediately.

CALCULATION


1. Subtract the absorbance of the blank from that of standards and samples.
2. Generate a standard curve by plotting the absorbance obtained (y-axis) against human cystatin c concentrations (x-axis). The best fit line can be generated with any curve-fitting software by regression analysis. Any curve of 4-parameter or log-log curve fitting can be used for calculation.
3. Determine human cystatin c concentration of samples from standard curve and multiply the value by the dilution factor.

TYPICAL STANDARD CURVE

The following standard curve is provided for demonstration only. A standard curve should be generated for each set of sample assay.

Cystatin c (ng/ml)	Absorbance (450 nm)	Blanked Absorbance
0	0.069	0
0.23	0.127	0.058
0.46	0.187	0.118
0.93	0.309	0.24
1.87	0.539	0.47
3.75	0.946	0.877
7.5	1.685	1.616
15	2.551	2.482

Human cystatin c standard curve (4-parameter)



ASSAY CHARACTERISTICS

A. Sensitivity: The lowest level of human cystatin c that can be detected by this assay is 0.23 ng/ml.

B. Specificity: The antibodies used in this assay are specific to human cystatin c and do not cross-react with mouse and rat cystatin c, and other cytokine or hormone molecules.

C. Precision:

Intra-assay Precision (Precision within an assay)

Three samples of known concentration were tested 12 times on one plate.

Sample	Mean (ng/ml)	SD (ng/ml)	CV (%)
1	507.3	30.1	5.9
2	377.2	23.1	6.1
3	137.4	10.8	7.9

Inter-assay Precision (Precision between assays)

Three samples of known concentration were tested in 10 separate assays.

Sample	Mean (ng/ml)	SD (ng/ml)	CV (%)
1	499.3	39.0	7.8
2	377.2	23.1	6.1
3	137.4	10.8	7.9

D. Recovery:

Serum samples were spiked with different amounts of human cystatin c and assayed.

Sample	Average % Recovery	Range %
Serum	99	94-107

REFERENCES

- [1] Stevens LA, Coresh J, Schmid CH, et al. (2008) *Am J Kidney Dis.* 51:395–406.
- [2] Dharnidharka VR, Kwon C, Stevens G. (2002) *Am. J. Kidney Dis.* 40 (2): 221–6.
- [3] Hermida J, Tutor JC. (2006) *Ther Drug Monit.* 28 (3): 326–31.
- [4] Zethelius B, Berglund L, Sundström J, et al. (2008) *N. Engl. J. Med.* 358 (20): 2107–16.
- [5] Ix JH, Shlipak MG, Chertow GM, Whooley MA. (2007) *Circulation* 115 (2): 173–9.

- [6] Deo R, Fyr CL, Fried LF, et al. (January 2008) *Am. Heart J.* 155 (1): 62–8.
[7] Servais A, Giral P, Bernard M, et al. (2008) *Am. J. Med.* 121 (5): 426–32.
[8] Mi W, Pawlik M, Sastre M, et al. (2007) *Nat. Genet.* 39 (12): 1440–2.
[9] Kaeser SA, Herzig MC, Coomaraswamy J, et al. (2007) *Nat. Genet.* 39 (12): 1437–9.
[10] Zurdel J, Finckh U, Menzer G, et al. (2002) *Br J Ophthalmol* 86 (2): 214–9.
[11] Strojan P, Oblak I, Svetic B, et al. (2004) *Br. J. Cancer* 90 (10): 1961–8.
[12] Kos J, Krasovec M, Cimerman N, et al. (2000) *Clin. Cancer Res.* 6 (2): 505–11.

SUMMARY OF ASSAY PROCEDURE

